

CERTIFICATE OF CALIBRATION

Issued By: PASS Ltd - www.calibrate.co.uk

Certificate Number: STD271018

Date of Issue: 29 Jul 2025

Page 1 of 3 Pages



PASS Ltd - www.calibrate.co.uk
1 Wilson Street
Thornaby
Teesside. TS17 7AR
TEL 01642 626144
calibrate@pass.co.uk

Signatory

A handwritten signature in black ink, appearing to read "P. Beswick".

Paul Beswick

Customer: That Event Company LTD
Unit 7 North Weston Farm
Thame
Oxfordshire
OX9 2HA

Date Received: 28 Jul 2025

Instrument:	System ID :	ID353043	Job Number :	J371737-1
	Description :	Multifunction Tester	Ref. Number :	
	Manufacturer :	Megger		
	Model Number :	MFT1741+		
	Serial Number :	102330613		
	Procedure :	4467 : 1.1.19		

Environmental Conditions:

Temperature :	20 °C ± 3 °C	Mains Voltage :	240 V ± 10 V
Relative Humidity :	50 %rh ± 20 %rh	Mains Frequency :	50 Hz ± 1 Hz

Comments: Specification reference: Megger (MFT1741_DS_en_V01) Instruction Manual.
Instrument was placed in lab and allowed to stabilise before calibration.
Instrument calibrated with in house test leads.

Traceability Information

Instrument Description	Serial Number	Certificate Number	Cal. Date	Cal. Period
3200B Electrical Test Calibrator	M1433L16	54891	21/07/2025	52

Calibrated By: R.Hussain

Date of Calibration: 29 Jul 2025

This certificate provides traceability of measurement to recognised National Standards, and to the units of measurements realised at the National Physical Laboratory or other recognised National Standards laboratories. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced except with the prior written approval of the issuing laboratory.

This certificate complies with the requirements of BS EN ISO 10012:2003.

CERTIFICATE OF CALIBRATION

Issued By: PASS Ltd - www.calibrate.co.uk
Date of Issue: 29 Jul 2025
AS FOUND RESULTS

Certificate Number
STD271018

Page 2 of 3 Pages

Test Title	Tolerance	Applied Value	Reading	Pass/Fail
General Operation Tests				
Controls & Display	---	---	Pass	Pass
Bar Display	---	---	Pass	Pass
Continuity Resistance				
0.5 Ω	30 m Ω	0.50 Ω	0.51 Ω	Pass
1 Ω	40 m Ω	1.00 Ω	1.01 Ω	Pass
10 Ω	400 m Ω	10.00 Ω	9.99 Ω	Pass
100 Ω	5.2 Ω	100.0 Ω	99.9 Ω	Pass
1.00 k Ω	70 Ω	1.00 k Ω	1.00 k Ω	Pass
Continuity Current @ 1Ω				
200 mA Test Current	15 mA	205.0 mA	216.6 mA	Pass
100V Insulation Range				
100 V Test @ 1 mA	20 V	100.0 V	109.8 V	Pass
1 M Ω	50 k Ω	1.00 M Ω	1.01 M Ω	Pass
9 M Ω	290 k Ω	9.00 M Ω	8.97 M Ω	Pass
20 M Ω	800 k Ω	20.0 M Ω	20.0 M Ω	Pass
180 M Ω	22 M Ω	180 M Ω	179 M Ω	Pass
250V Insulation Range				
250 V Test @ 1 mA	50 V	250.0 V	268.6 V	Pass
1 M Ω	50 k Ω	1.00 M Ω	1.00 M Ω	Pass
9 M Ω	290 k Ω	9.00 M Ω	8.96 M Ω	Pass
20 M Ω	800 k Ω	20.0 M Ω	20.0 M Ω	Pass
290 M Ω	33 M Ω	290 M Ω	289 M Ω	Pass
500V Insulation Range				
500 V Test @ 1 mA	100 V	500.0 V	543.1 V	Pass
1 M Ω	50 k Ω	1.00 M Ω	1.00 M Ω	Pass
9 M Ω	290 k Ω	9.00 M Ω	8.96 M Ω	Pass
20 M Ω	800 k Ω	20.0 M Ω	20.0 M Ω	Pass
290 M Ω	10.7 M Ω	290 M Ω	289 M Ω	Pass
590 M Ω	63 M Ω	590 M Ω	579 M Ω	Pass
1kV Insulation Range				
1 kV Test @ 1 mA	200 V	1000.0 V	1026.9 V	Pass
1 M Ω	50 k Ω	1.00 M Ω	1.01 M Ω	Pass
9 M Ω	290 k Ω	9.00 M Ω	8.99 M Ω	Pass
20 M Ω	800 k Ω	20.0 M Ω	20.0 M Ω	Pass
290 M Ω	10.7 M Ω	290 M Ω	291 M Ω	Pass
590 M Ω	19.7 M Ω	590 M Ω	593 M Ω	Pass

CERTIFICATE OF CALIBRATION

Issued By: PASS Ltd - www.calibrate.co.uk
Date of Issue: 29 Jul 2025
AS FOUND RESULTS

Certificate Number
STD271018

Page 3 of 3 Pages

Test Title	Tolerance	Applied Value	Reading	Pass/Fail
Voltage Tests				
	10.1 V	237 V	237 V	Pass
	14.8 V	395 V	395 V	Pass
Loop Test 25A Range				
0.5 Ω	76.3 mΩ	0.53 Ω	0.54 Ω	Pass
1 Ω	100.8 mΩ	1.02 Ω	0.99 Ω	Pass
5 Ω	326.8 mΩ	5.54 Ω	5.57 Ω	Pass
10 Ω	977.4 mΩ	9.5 Ω	9.6 Ω	Pass
100 Ω	14.1 Ω	91 Ω	92 Ω	Pass
1000 Ω	104.4 Ω	994 Ω	999 Ω	Pass
Loop Test No Trip Range				
0.5 Ω	76.3 mΩ	0.53 Ω	0.55 Ω	Pass
1 Ω	100.8 mΩ	1.02 Ω	1.03 Ω	Pass
10 Ω	977.4 mΩ	9.5 Ω	9.5 Ω	Pass
Trip Current @ 200ms				
10 mA	1 mA	10 mA	10.75 mA	Pass
30 mA	3 mA	30 mA	31.89 mA	Pass
100 mA	10 mA	100 mA	105.28 mA	Pass
300 mA	30 mA	300 mA	316.34 mA	Pass
500 mA	50 mA	500 mA	522.91 mA	Pass
I Multiplier @ 200ms				
30 mA @ 5 I	15 mA	150.00 mA	157.73 mA	Pass
Trip Time				
30 ms @ 30 mA	1.3 ms	30.0 ms	30.8 ms	Pass
90 ms @ 30 mA	1.9 ms	90.0 ms	90.9 ms	Pass
290 ms @ 30 mA	3.9 ms	290 ms	290 ms	Pass
Earth Resistance 3P				
1 Ω	50 mΩ	1.00 Ω	0.99 Ω	Pass
10 Ω	230 mΩ	10.00 Ω	9.93 Ω	Pass
100 Ω	2.3 Ω	100.0 Ω	99.6 Ω	Pass
1000 Ω	50 Ω	1.00 kΩ	1.01 kΩ	Pass

**** END OF RESULTS ****

Uncertainties

- Loop Impedance 1 Ω ± 8.6 mΩ; 10 Ω ± 30 mΩ; 100 Ω ± 0.57 Ω; 1 kΩ ± 5.8 Ω ±1 LSD
- Continuity Ohms 20 mΩ ± 30 mΩ; 1 Ω ± 32 mΩ; 10 Ω ± .12 Ω; 100 Ω ± 1.2 Ω; 1 kΩ ± 12 Ω ±1 LSD
- Continuity Current 100 mA ± 1.7 mA; 200 mA ± 3.1 mA; 300 mA ± 4.6 mA ±1 LSD
- Insulation Ohms 0.1 MΩ ± 0.35 kΩ; 1 MΩ ± 3.5 kΩ; 10 MΩ ± 0.35 MΩ; 100 MΩ ± 3.5 MΩ; 1 GΩ ± 36 MΩ ±1 LSD
- Insulation Voltage 100 V ± 1.5 V; 150 V ± 2.1 V; 250 V ± 3.1 V; 500 V ± 5.9 V; 1000 V ± 12 V ±1 LSD
- RCD Current 10 mA ± 0.62 mA; 30 mA ± 1.8 mA; 100 mA ± 5.8 mA; 300 mA ± 18 mA; 1 A ± 58 mA ±1 LSD
- RCD Time 20 to 390 ms ± 0.95 ms; 900 ms ± 8.1 ms
- Mains Voltage 200 V to 260 V ± 2.4 V ±1 LSD