

PB500 ANNUAL CERTIFICATE OF CALIBRATION

Customer	PAC Serial No	Expiry Date	Certificate No
That Event Company Ltd	25E-0576	18-Feb-23	60046118

The measurements reported in this certificate were carried out using equipment whose measured value are traceable to National Standards, where these exist.

The ambient temperature throughout the calibration was 18-22 degrees.

UNIT UNDER TEST	PB500
TEST STATUS	Complete and correct

<u>Earth Bond Test</u>	<u>Actual Value (Ohm)</u>
------------------------	---------------------------

PASS	0.25 ohms
-------------	------------------

FAIL	0.35 ohms
-------------	------------------

<u>Insulation Resistance</u>	<u>Actual Value (Mohm)</u>
------------------------------	----------------------------

PASS	2.2 Mohms
-------------	------------------

FAIL	1.8 Mohms
-------------	------------------

DC VOLTAGE	PASS (450 V) Minimum
-------------------	-----------------------------

Electrical safety tests were carried out and found to be satisfactory.

Equipment Used for Calibration

PAC500 Test Rig	Serial No: K02-0184	Certificate No: SEA045049/2
------------------------	----------------------------	------------------------------------

PB500 Test Rig	Serial No: TS957	Certificate No: SEA048978/2R
-----------------------	-------------------------	-------------------------------------

A record of this certificate is kept at Parker Bell Instruments Ltd